

Syddansk Universitet

Adapting bilateral directional processing to individual and situational influences

Neher, Tobias; Wagener, Kirsten C.; Latzel, Matthias

Publication date:
2017

Document version
Publisher's PDF, also known as Version of record

Citation for pulished version (APA):
Neher, T., Wagener, K. C., & Latzel, M. (2017). Adapting bilateral directional processing to individual and situational influences. Poster session presented at International Symposium on Auditory and Audiological Research 2017, Nyborg, Denmark.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal ?

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Tobias Neher¹, Kirsten C. Wagener² & Matthias Latzel³

¹ Medizinische Physik, Oldenburg University; ² Hörzentrum Oldenburg GmbH; ³ Phonak AG, Switzerland


INTRODUCTION

Hearing aid (HA) users can differ markedly in their benefit from directional processing algorithms. The current study (Neher et al, 2017) thus investigated candidacy for different directional processing schemes. In a previous study (Neher, 2017), we screened ~80 elderly hearing-impaired listeners with a large spread in the degree of audiometric asymmetry <2 kHz (Δ PTALF) re. the binaural contribution to speech-in-noise reception (BILD). Here, we tested a subset of these listeners using computer simulations of different HA fittings and acoustic scenarios. Our **aims** were to:

- (1) Relate Δ PTALF and BILD to performance with five directional processing schemes that differed in the trade-off between signal-to-noise ratio (SNR) improvement and binaural cue preservation
- (2) Investigate if a simple binaural tone-in-noise detection measure can predict benefit from binaural cue preservation



PARTICIPANTS

- Groups of listeners with symmetric ($N = 20$) or asymmetric ($N = 19$) Δ PTALF, a large spread in the binaural intelligibility level difference (BILD; Kollmeier, 1996), and no difference in age, overall degree of hearing loss (PTA4), N_0S_{π} detection performance at 500 Hz or distractibility/selective attention (cf. Neher, 2017)





STIMULI

- Virtual acoustics: Head-related impulse responses measured with two behind-the-ear (BTE) HA dummies placed on a head-and-torso simulator (Kayser et al, 2009)
- Target: OISa sentences (Wagener et al, 1999) from 0° and 1 m
- Maskers:
 - olsa60**: OISa sentences from $\pm 60^\circ$ (free-field, different speaker)
 - ists60**: International Speech Test Signal (Holube et al, 2010) from $\pm 60^\circ$ (free-field, different speakers)
 - cafnois**: Spatially diffuse speech babble with intermittent voices and other noises



HA CONDITIONS

- Simulation of a linked pair of completely occluding BTE devices based on the Master Hearing Aid (Grimm et al, 2006); full exchange of audio signals
 - pinna**: Slightly directional >2 kHz (dichotic)
 - beamfull**: Highly directional re. 0° (diotic)
 - beam>0.8k**: beamfull >0.8 kHz, pinna below (hybrid)
 - beam<2k**: beamfull <2 kHz, pinna above (hybrid)
 - beambetter**: beamfull with only better ear stimulated (monaural)




- Speech-weighted SNR improvement (Δ AI-SNR) re. pinna, see table \rightarrow
- Amplification according to NAL-RP (Dillon, 2012)

HA condition	Δ AI-SNR re. pinna (dB)			
	olsa60	ists60	cafnois	mean
beamfull	4.2	4.8	4.0	4.4
beam>0.8k	2.0	2.0	2.3	2.1
beam<2k	3.2	3.6	2.6	3.1
beambetter	4.2	4.8	4.0	4.4


RESULTS

- Speech reception thresholds (SRTs) corresponding to 50%-correct speech intelligibility
- Very good test-retest reliability (all $r > 0.73$, all $p < 0.00001$)
- Δ PTALF affects influence of HA condition in olsa60 and ists60 scenarios



Symmetric group

- Repeated-measures ANOVA, main finding:
 - BILD \times HA condition \times acoustic scenario: $F_{(3.3, 53.1)} = 6.3$, $p = 0.0006$



Asymmetric group

- Repeated-measures ANOVA, main finding:
 - BILD \times HA condition \times acoustic scenario: $F_{(3.7, 55.4)} = 3.4$, $p = 0.016$





BILD vs. N_0S_{π}

- BILD and N_0S_{π} strongly correlated ($r = -0.72$, $p < 0.00001$)
- Repeated-measures ANOVA with N_0S_{π} instead of BILD (data of symmetric group)
 - $N_0S_{\pi} \times$ HA condition \times acoustic scenario: $F_{(3.2, 50.7)} = 4.2$, $p = 0.009$



SUMMARY & OUTLOOK

- (1) Binaural hearing abilities, audiometric asymmetry <2 kHz and the acoustic scenario influence speech reception with bilateral directional processing
- (2) For lateral speech maskers, binaural hearing abilities modulate benefit from preserved low-frequency binaural cues
- (3) For spatially diffuse noise, the maximal SNR improvement is beneficial
- (4) Audiometric asymmetry <2 kHz reduces the influence of binaural hearing
- (5) N_0S_{π} detection at 500 Hz predicts benefit from low-frequency binaural cues effectively

These findings provide a basis for adapting directional processing to the user and the scenario. Ongoing research investigates their generalizability to clinical HA fittings.

Acknowledgements: Funded by the DFG Cluster of Excellence EXC 1077/1 "Hearing4all" and by Sonova AG, Switzerland

REFERENCES

- Dillon H (2012) *Hearing Aids*, 2nd ed. Boomerang Press, Sydney, Australia
- Grimm G et al (2006) "The master hearing aid: A PC-based platform for algorithm development and evaluation," *Acta Acust Acust*, 92:618-628
- Holube I et al (2010) "Development and analysis of an International Speech Test Signal (ISTS)," *Int J Audiol*, 49: 891-903
- Kayser H et al (2009) "Database of multichannel in-ear and behind-the-ear head-related and binaural room impulse responses," *EURASIP J Adv Signal Process*, DOI: 10.1155/2009/298605
- Kollmeier B (1996) "Computer-controlled speech audiometric techniques for the assessment of hearing loss and the evaluation of hearing aids," In: Kollmeier B (Ed.), *Psychoacoustics, Speech and Hearing Aids*. World Scientific Publishing, Singapore, 57-68
- Neher T (2017) "Characterizing the binaural contribution to speech-in-noise reception in elderly hearing-impaired listeners," *J Acoust Soc Am*, 141, EL159-EL163
- Neher T et al (2017) "Speech reception with different bilateral directional processing schemes: Influence of binaural hearing, audiometric asymmetry, and acoustic scenario," *Hear Res*, 353, 36-48
- Wagener KC et al (1999) "Entwicklung und Evaluation eines Satztests für die deutsche Sprache, I-III: Design, Optimierung und Evaluation des Oldenburger Satztests", *Zeitschrift für Audiologie (Audiological Acoustics)*, 38, 4-15, 44-56, 86-95